



Attestation of Conformity

Attestation No. 16411

Certificate holder: Lenard Enterprises Inc.
Address: 1-1211 Gorham St., Newmarket ON L3Y8Y3 Canada.
Product Type: Smart Ultrasound transducer Analyzer
Model No. : ELZHEN 2200

The equipment complies with the principal requirements of the Council Directive 2014/35/EU
Low Voltage Directive

The application of the CE mark is permitted after all applicable requirements are met in accordance with the
European Union rules.

This attestation applies to the particular sample of the product provided for testing and certification.
The detailed test results and standards used as well as the operation mode are listed in

Test report number: 1749-81400-D
Date of issuance: 15.11.2021
**Applied specifications/
standards:** IEC61010-1(2017)
Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory
Use – Part 1: General Requirements.
IEC61326-1:2020
Electrical Equipment for Measurement, Control and Laboratory Use – EMC Requirements
– Part 1: General Requirements.

After preparation of the necessary technical documentation as well as the conformity declaration the CE marking as shown can be
affixed on the equipment. Other relevant directives have to be observed

For precise and updated information concerning
possible changes occurred in the certification object of
the present certificate, please contact
info@quality-service.cn

Audit date: 04.04.2022
Date of issue: 07.04.2022
Expiration date: 06.04.2025

QS Zürich AG
P.O. Box 6335
CH-8050 Zürich
qs-zuerich@quality-service.ch



A blue ink signature.

Direction